

ELISA Technologies, Inc.



elisa-tek.com

PAGE 1 OF 1 - REPORT 2207101Y (of A-AH)

REPORT DATE: July 26, 2022

REPORT TO: Eva Lau

Mayway Corp.

1338 Mandela Parkway Oakland, CA 94607 Tel: 510-208-3113

Email: evalau@mayway.com

SUBJECT: PO Number: 071422; Samples for Gluten Analysis

RECEIVED: July 20, 2022; UPS Ground TRK# 1Z 89E 637 03 6166 8349

ANALYTICAL RESULTS: RESULTS APPLY ONLY TO THE SAMPLES AS RECEIVED.

I	CONTROL #	SAMPLE DESCRIPTION	ANALYSIS IDENTIFICATION	Gluten
	2207101-25	Item# 3366E Product: Four Gentlemen Teapills Si Jun Zi Wan Batch: A9H1-1	500821 Quantitative Gluten ELISA	> 80.0 ppm

NOTE: The ALLER-TEK® Gluten assay is a sandwich ELISA for the quantitative analysis of gluten in raw materials and finished food products. Results are expressed in ppm gluten with a quantifiable range of 2.5-80 ppm; it detects the gliadin and glutenin fractions of gluten. A result of "not detected" indicates that gluten was not detected in the sample (<2.5 ppm). Recovery may be reduced in samples that have been hydrolyzed.

The ALLER-TEK $^{\circ}$ Gluten ELISA has been validated and certified with license no. 081202 as a Performance Tested Method by the AOAC Research Institute.

Information in italics has been supplied by the customer.

Respectfully Submitted for ELISA Technologies, Inc.

By James Tom d'angelo

By ______Justin Bickford, Scientific Director

End of Report

THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN ITS ENTIRETY AND WITH THE WRITTEN PERMISSION OF ELISA TECHNOLOGIES, INC.