



# ELISA Technologies, Inc.

**elisa-tek.com**



PAGE 1 OF 1 - REPORT 2208090H (of A-K)

REPORT DATE: August 18, 2022

REPORT TO: *Eva Lau*  
*Mayway Herbs*  
*1338 Mandela Parkway*  
*Oakland, CA 94607*  
*Tel: 510-208-3113*  
*Email: evalau@mayway.com*

SUBJECT: *PO Number: GLU080422; Sample for Gluten Analysis*

RECEIVED: *August 12, 2022; UPS Ground TRK# 1Z 89E 637 03 6125 7319*

**ANALYTICAL RESULTS: RESULTS APPLY ONLY TO THE SAMPLES AS RECEIVED.**

| CONTROL # | SAMPLE DESCRIPTION                                                                      | ANALYSIS IDENTIFICATION                     | Gluten               |
|-----------|-----------------------------------------------------------------------------------------|---------------------------------------------|----------------------|
| 2208090-8 | <i>Item# 3625C Product: Si Miao San,<br/>extract powder Si Miao San<br/>Batch: B0E1</i> | <i>500821 Quantitative<br/>Gluten ELISA</i> | <b>&gt; 80.0 ppm</b> |

**NOTE:** The ALLER-TEK® Gluten assay is a sandwich ELISA for the quantitative analysis of gluten in raw materials and finished food products. Results are expressed in ppm gluten with a quantifiable range of 2.5-80 ppm; it detects the gliadin and glutenin fractions of gluten. A result of "not detected" indicates that gluten was not detected in the sample (<2.5 ppm). Recovery may be reduced in samples that have been hydrolyzed.

The ALLER-TEK® Gluten ELISA has been validated and certified with license no. 081202 as a Performance Tested Method by the AOAC Research Institute.

Information in italics has been supplied by the customer.

**Respectfully Submitted for ELISA Technologies, Inc.**

By *Tom d'Angelo*  
Analyst

By *Kerim Gattas Asfura*  
Kerim Gattas Asfura, Testing Supervisor

End of Report

THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN ITS ENTIRETY AND WITH THE WRITTEN PERMISSION OF ELISA TECHNOLOGIES, INC.

2501 NW 66<sup>th</sup> Court, Gainesville, Florida 32653, USA  
+1 352 337 3929 info@elisa-tek.com